

Certificate of Analysis

Product name	AMPLIRUN® STREPTOCOCCUS AGALACTIAE DNA CONTROL
Catalogue Number	MBC071-R
Lot Number	25MBC071001-R
Expiry date	2027-11-30

Kit contents	Lot Number	Expiry date
VIRCELL SAG DNA CONTROL	25MBC071101-R01	2027-11-30

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	11203 ± 672
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

dltS.F CTGTAAGTCTTTATCTTTCTCG
dltS.R TCCATTCGCTTAGTCTCC

Morozumi, M., et al. (2015). J Infect Chemother, 21(1):34-8

VIRCELL CONTROL RECONSTITUTION SOLUTION	25MBCRS001-R	2030-01-20
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Tejera García, Hugo	09/06/2025	