

Certificate of Analysis

Product name AMPLIRUN® ZIKA VIRUS RNA CONTROL
Catalogue Number MBC072-R
Lot Number 25MBC072001-R
Expiry date 2026-07-08

Kit contents	Lot Number	Expiry date
VIRCELL ZIKV RNA CONTROL	25MBC072101-R01	2026-07-08

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	17247 ± 804
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides
 ZIKW-F CAGCTGGCATCATGAAGAAYC
 ZIKW-R CACTTGTCCTCTTCTCTCC
 Waggoner, J. J., et al. (2016). J Clin Virol, 78(5):57-61

VIRCELL CONTROL RECONSTITUTION SOLUTION	25MBCRS001-R	2030-01-20
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Tejera García, Hugo	17/06/2025	