

Certificate of Analysis

Product name AMPLIRUN® ZIKA VIRUS RNA CONTROL
Catalogue Number MBC072-R
Lot Number 24MBC072001-R
Expiry date 2026-07-08

Kit contents	Lot Number	Expiry date
VIRCELL ZIKV RNA CONTROL	24MBC072101-R01	2026-07-08

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	17247 ± 804
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION
Identity test oligonucleotides ZIKW-F CAGCTGGCATCATGAAGAAYC ZIKW-R CACTTGTCCTCTTCTCTCC <i>Waggoner, J. J., et al. (2016). J Clin Virol, 78(5):57-61</i>		

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by

Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	16/07/2024	