

Certificate of Analysis

Product name	AMPLIRUN® NEISSERIA GONORRHOEAE DNA CONTROL
Catalogue Number	MBC075
Lot Number	24MBC075001-C
Expiry date	2025-09-28

Kit contents	Lot Number	Expiry date
VIRCELL NGO DNA CONTROL	24MBC075101-C01	2025-09-28

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	16000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides NGO-16S.F TATCGGAACGTACCGGGTAGC NGO-16S.R GTATTACCGCGGCTGCTGGCA <i>Chui, L., et al. (2008). Clin Microbiol Infect, 14(5):473-479</i>		
---	--	--

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-C	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE and part of Regulation (EU) 2017/746. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	28/06/2024	