

# Certificate of Analysis

Product name	AMPLIRUN® BORRELIA BURGdorferi DNA CONTROL
Catalogue Number	MBC076-R
Lot Number	23MBC076001-R
Expiry date	2025-12-13


Kit contents	Lot Number	Expiry date
VIRCELL BBU DNA CONTROL	23MBC076101-R01	2025-12-13

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	16000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

<b>Identity test oligonucleotides</b> <i>B.burdo.F</i> CTGTTGGTGAGCTTCTTACT <i>B.burdo.R</i> GCATTATACTCATTATGGTAGA  <i>Lee, S.H., et al. (2000). J Clin Microbiol, 38(7):2557-2562</i>		
---	--	--

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Reca John, Isabel María	23/06/2023	



Vircell, S.L.  
 Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264  
 info@vircell.com www.vircell.com