

Certificate of Analysis

Product name	AMPLIRUN® BORRELIA GARINII DNA CONTROL
Catalogue Number	MBC077-R
Lot Number	24MBC077001-R
Expiry date	2026-12-26

Kit contents	Lot Number	Expiry date
VIRCELL BGA DNA CONTROL	24MBC077101-R01	2026-12-26

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	16413 ± 203
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION
Identity test oligonucleotides <i>B.garinii.F</i> AGTCCCCCTGGTCCAAGG <i>B.garinii.R</i> GTGCGTTCTGTTGGGGAG Lee, S.H., et al. (2000). J Clin Microbiol, 38(7):2557-2562		

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by

Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	02/07/2024	