

Certificate of Analysis

Product name	AMPLIRUN® BORRELIA AFZELII DNA CONTROL
Catalogue Number	MBC078-R
Lot Number	25MBC078001-R
Expiry date	2027-08-25

Kit contents	Lot Number	Expiry date
VIRCELL BAZ DNA CONTROL	25MBC078101-R01	2027-08-25

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	14045 ± 204
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

Bo_af_flg_F GGAGCAAATCAAGATGAAGCAAT

Bo_af_flg_R TGAGCACCTCTTGAACAGG

Raileanu, C., et al. (2017). Front Cell Infect Microbiol, 7:36

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Tejera García, Hugo	03/03/2025	