

Certificate of Analysis

Product name	AMPLIRUN® TRICHOMONAS VAGINALIS DNA CONTROL
Catalogue Number	MBC079
Lot Number	25MBC079001-C
Expiry date	2027-06-23

Kit contents	Lot Number	Expiry date
VIRCELL TVA DNA CONTROL	25MBC079101-C01	2027-06-23

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	17700 ± 1080
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

TVA-F CATTGATAACGAAGCTCTTTACGAT

TVA-R GCATGTTGTGCCGGACATAACCAT

Hardick, J., et al. (2003). J Clin Microbiol, 41(12):5619-5622

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-C	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the applicable general safety and performance requirements as established in Regulation (EU) 2017/746. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Tejera García, Hugo	20/01/2025	