

# Certificate of Analysis

Product name	AMPLIRUN® NOVEL INFLUENZA A H1N1 RNA CONTROL
Catalogue Number	MBC082
Lot Number	24MBC082001-C
Expiry date	2026-07-23

Kit contents	Lot Number	Expiry date
VIRCELL INAH1N1 RNA CONTROL	24MBC082101-C01	2026-07-23

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	18188 ± 1242
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

## Identity test oligonucleotides

Flu A Fw GACCRATCCTGTACCTCTGAC  
Flu A Rv GGGCATTYTGACAAAKCGTCTACG

Chen, Y., et al. (2011). J Clin Microbiol, 49(4):1653-1656

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-C	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE and part of Regulation (EU) 2017/746. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

## Q.C. Testing by

Position	Name	Date	Signature
Quality Control Department	Martínez Ramírez, David	01/08/2024	