

Certificate of Analysis

Product name	AMPLIRUN® MYCOPLASMA GENITALIUM DNA CONTROL
Catalogue Number	MBC085
Lot Number	23MBC085001-C
Expiry date	2025-08-14

Kit contents	Lot Number	Expiry date
VIRCELL MGE DNA CONTROL	23MBC085101-C01	2025-08-14

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	13000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides


MGE-Mg219-F CATAGTTCATTATGCGCACCAGTTACTTG
MGE-Mg219-R CTCTTTAACACAGGGGTTGGGATTAG

Chalker, V.J., et al. (2009). J Med Microbiol., 58(7):895-899

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS001-C	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE.

Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Guerrero Ruiz, Juan José	22/02/2023	



Vircell, S.L.
Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264
info@vircell.com www.vircell.com