

Certificate of Analysis

Product name	AMPLIRUN® MYCOPLASMA GENITALIUM DNA CONTROL
Catalogue Number	MBC085
Lot Number	24MBC085002-C
Expiry date	2027-01-16

Kit contents	Lot Number	Expiry date
VIRCELL MGE DNA CONTROL	24MBC085102-C01	2027-01-16

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	16140 ± 207
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

MGE-Mg219-F CATAGTTCATTATGCGCACCAGTTACTTG
MGE-Mg219-R CTCTTAACAACAGGGGTGGGATTAG

Chalker, V.J., et al. (2009). J Med Microbiol., 58(7):895-899

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-C	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE and part of Regulation (EU) 2017/746. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by

Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	24/07/2024	