

# Certificate of Analysis

Product name	AMPLIRUN® MYCOBACTERIUM AVIUM DNA CONTROL
Catalogue Number	MBC086-R
Lot Number	24MBC086001-R
Expiry date	2027-03-05

Kit contents	Lot Number	Expiry date
VIRCELL MAV DNA CONTROL	24MBC086101-R01	2027-03-05

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	17540 ± 614
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**

Mav-IS1311-f GCTGGACGCATTACGCAAT

Mav-IS1311-r CGTCAGCTCCGCATCGAT

Sevilla, I. A., et al. (2015). J Clin Microbiol, 53(3):930-940

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.  
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	10/09/2024	