

Certificate of Analysis

Product name	AMPLIRUN® MYCOBACTERIUM INTRACELLULARE DNA CONTROL
Catalogue Number	MBC087-R
Lot Number	23MBC087001-R
Expiry date	2025-08-20

Kit contents	Lot Number	Expiry date
VIRCELL MIN DNA CONTROL	23MBC087101-R01	2025-08-20

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	11000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION


Identity test oligonucleotides

MIN-DT1-F AAGGTGAGCCCAGCTTTGAACTCCA
MIN-DT1-R GCGCTTCATTCGCGATCATCAGGTG

Chae, H., et al. (2017). J Clin Microbiol, 55(9):2736-2751

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Guerrero Ruiz, Juan José	01/03/2023	



Vircell, S.L.
Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264
info@vircell.com www.vircell.com