

# Certificate of Analysis

Product name	AMPLIRUN® MYCOBACTERIUM INTRACELLULARE DNA CONTROL
Catalogue Number	MBC087-R
Lot Number	25MBC087001-R
Expiry date	2027-09-19

Kit contents	Lot Number	Expiry date
VIRCELL MIN DNA CONTROL	25MBC087101-R01	2027-09-19

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	17128 ± 199
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**

MIN-DT1-F AAGGTGAGCCCAGCTTTGAACTCCA  
 MIN-DT1-R GCGCTTCATTCGCGATCATCAGGTG

Chae, H., et al. (2017). J Clin Microbiol, 55(9):2736-2751

VIRCELL CONTROL RECONSTITUTION SOLUTION	25MBCRS001-R	2030-01-20
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.  
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	28/03/2025	