

Certificate of Analysis

Product name	AMPLIRUN® CAMPYLOBACTER JEJUNI DNA CONTROL
Catalogue Number	MBC088-R
Lot Number	24MBC088001-R
Expiry date	2026-09-18

Kit contents	Lot Number	Expiry date
VIRCELL CJE DNA CONTROL	24MBC088101-R01	2026-09-18

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	15000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

CJE.VS15 GAATGAAATTTTGAATGGGG
CJE.VS16 GATATGTATGATTTATCCTGC

Yang, C., et al. (2003). FEMS Immunol Med Microbiol, 38(3):265-271

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-R	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	26/03/2024	