

# Certificate of Analysis

Product name	AMPLIRUN® CORONAVIRUS RNA CONTROL
Catalogue Number	MBC090-R
Lot Number	24MBC090001-R
Expiry date	2026-08-26

Kit contents	Lot Number	Expiry date
VIRCELL COR RNA CONTROL	24MBC090101-R01	2026-08-26

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	14090 ± 240
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**

CoV229E-F CAGTCAAATGGGCTGATGCA

CoV229E-R AAAGGGCTATAAGAGAATAAGGTATTCT

Zhang, S.F., et al. (2018).PLOS ONE, 13(1):e0191789

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.  
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	06/09/2024	