

Certificate of Analysis

Product name	AMPLIRUN® TRYPANOSOMA RANGELI DNA CONTROL
Catalogue Number	MBC093-R
Lot Number	23MBC093001-R
Expiry date	2025-11-15


Kit contents	Lot Number	Expiry date
VIRCELL TRA DNA CONTROL	23MBC093101-R01	2025-11-15

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	15000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides <i>TrF3</i> CCCATACAAAACACCCCTT <i>TrR8</i> TGGAATGACGGTGCGGCGAC <i>Chiurillo, M.A., et al. (2003). Clin Diagn Lab Immunol. 10(5):775-77</i>		
---	--	--

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Reca John, Isabel María	23/06/2023	



Vircell, S.L.
 Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264
 info@vircell.com www.vircell.com