

Certificate of Analysis

Product name	AMPLIRUN® VENEZUELAN EQUINE ENCEPHALITIS VIRUS RNA CONTROL
Catalogue Number	MBC096-R
Lot Number	23MBC096001-R
Expiry date	2025-07-06

Kit contents	Lot Number	Expiry date
VIRCELL VEEV RNA CONTROL	23MBC096101-R01	2025-07-06


Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	15000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides
 VEEZ-forward CATAGTCTAGTCGCGCAAGATGTT
 VEEZ-reverse CGATAGGGCATTGGCTGCAT

Smith, D. R., et al. (2006). *Emerg Infect Dis*, 12(8):1190-1196

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	27/07/2023	



Vircell, S.L.
 Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264
 info@vircell.com www.vircell.com