

Certificate of Analysis

Product name	AMPLIRUN® CHIKUNGUNYA VIRUS RNA CONTROL
Catalogue Number	MBC099-R
Lot Number	25MBC099001-R
Expiry date	2027-06-03

Kit contents	Lot Number	Expiry date
VIRCELL CHIK RNA CONTROL	25MBC099101-R01	2027-06-03

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	15563 ± 108
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides CHKW-F CATCTGCACYCAAGTGACCA CHKW-R GCGCATTTTGCCTTCGTAATG <i>Waggoner, J. J., et al.(2016). J Clin Virol, 78:57-61</i>		
--	--	--

VIRCELL CONTROL RECONSTITUTION SOLUTION	25MBCRS001-R	2030-01-20
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Tejera García, Hugo	09/06/2025	