

Certificate of Analysis

Product name	AMPLIRUN® YELLOW FEVER VIRUS RNA CONTROL
Catalogue Number	MBC100-R
Lot Number	24MBC100001-R
Expiry date	2026-06-13

Kit contents	Lot Number	Expiry date
VIRCELL YFEV RNA CONTROL	24MBC100101-R01	2026-06-13

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	13289 ± 261
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION
Identity test oligonucleotides YFVNS5-F: CATGGTCGATTCATGGGAAAG YFVNS5-R: CGCACAGCTTGCTTGCTC <i>Nunes, M. R. T., et al. (2011). J Virol Methods, 174(1-2):29-34</i>		

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	21/06/2024	