

Certificate of Analysis

Product name	AMPLIRUN® ST LOUIS ENCEPHALITIS VIRUS RNA CONTROL
Catalogue Number	MBC101-R
Lot Number	24MBC101002-R
Expiry date	2026-06-25

Kit contents	Lot Number	Expiry date
VIRCELL STLE RNA CONTROL	24MBC101102-R01	2026-06-25

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	14375 ± 347
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

SLE2420 CTGGCTGTCTGGAGGGATTCT
SLE2487c TAGGTCAATTGCACATCCCG

Lanciotti, R. S. & Kerst, A. J. (2001). J Clin Microbiol, 39(12):4506-4513

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	28/11/2024	