

Certificate of Analysis

Product name	AMPLIRUN® ZIKA VIRUS (ASIAN LINEAGE) RNA CONTROL
Catalogue Number	MBC103-R
Lot Number	24MBC103001-R
Expiry date	2026-10-10

Kit contents	Lot Number	Expiry date
VIRCELL ZIKV (ASIAN LINEAGE) RNA CONTROL	24MBC103101-R01	2026-10-10

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	13283 ± 344
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

ZIKW-F CAGCTGGCATCATGAAGAAYC

ZIKW-R CACTTGTCCCATCTTCTCTCC

Waggoner, J. J., et al. (2016). J Clin Virol, 78(5):57-61

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	19/11/2024	