

Certificate of Analysis

Product name	AMPLIRUN® HUMAN PARAINFLUENZA 1 RNA CONTROL
Catalogue Number	MBC105-R
Lot Number	24MBC105001-R
Expiry date	2026-05-10

Kit contents	Lot Number	Expiry date
VIRCELL HPAR1 RNA CONTROL	24MBC105101-R01	2026-05-10

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	13417 ± 898
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION
Identity test oligonucleotides PIS1+ CCGGTAATTTCTCATACCTATG PIS1- CTTTGGAGCGGAGTTGTTAAG <i>Echevarria, J. E., et al. (1998). J Clin Microbiol, 36(5):1388-1391</i>		

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-R	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by

Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	23/05/2024	



Vircell, S.L.
 Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264
 info@vircell.com www.vircell.com