

Certificate of Analysis

Product name AMPLIRUN® NOROVIRUS RNA CONTROL
Catalogue Number MBC111-R
Lot Number 24MBC111001-R
Expiry date 2026-09-12

Kit contents	Lot Number	Expiry date
VIRCELL NOR RNA CONTROL	24MBC111101-R01	2026-09-12

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	700 - 2000	1511 ± 82
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:100 DILUTION	SPECIFIC AMPLIFICATION OF 1:100 DILUTION

Identity test oligonucleotides
NoroCOG2F CARGARBCNATGTTYAGRTGGATGAG
NoroCOG2R TCGACGCCATCTTCATTCACA
Kageyama, T., et al. (2003). J Clin Microbiol, 41(4):1548-1557

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	25/09/2024	