

Certificate of Analysis

Product name AMPLIRUN® ADENOVIRUS 41 DNA CONTROL
Catalogue Number MBC114-R
Lot Number 25MBC114001-R
Expiry date 2027-09-13

Kit contents	Lot Number	Expiry date
VIRCELL ADN41 DNA CONTROL	25MBC114101-R01	2027-09-13

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	15668 ± 539
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

JTVFF (619-641) AACTTTCTCTCTTAATAGACGCC
JTVFR (736-718) AGGGGGCTAGAAAACAAAA

Jothikumar, N., et al. (2005). Appl Environ Microbiol, 71(6):3131-3136

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	21/03/2025	