

Certificate of Analysis

Product name	AMPLIRUN® VIBRIO CHOLERAЕ DNA CONTROL
Catalogue Number	MBC118-R
Lot Number	24MBC118001-R
Expiry date	2026-11-14

Kit contents	Lot Number	Expiry date
VIRCELL VCH DNA CONTROL	24MBC118101-R01	2026-11-14

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	15715 ± 318
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION
Identity test oligonucleotides VCHO-rtxA-V1F AGCAAGAGCATTGTTGTTCTACC VCHO-rtxA-V1R ACTTCCCTGTACCGCACTTAGAC Gubala, A.J. (2006) J Microbiol Methods, 65(2):278-293		

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-R	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by

Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	27/05/2024	