

Certificate of Analysis

Product name	AMPLIRUN® VIBRIO CHOLERAЕ DNA CONTROL
Catalogue Number	MBC118-R
Lot Number	25MBC118001-R
Expiry date	2026-11-14

Kit contents	Lot Number	Expiry date
VIRCELL VCH DNA CONTROL	25MBC118101-R01	2026-11-14

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	15715 ± 318
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

VCHO-rtxA-V1F AGCAAGAGCATTGTTGTTCTACC
VCHO-rtxA-V1R ACTTCCCTGTACCGCACTTAGAC

Gubala, A.J. (2006) J Microbiol Methods, 65(2):278-293

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	07/04/2025	