

Certificate of Analysis

Product name AMPLIRUN® GIARDIA INTESTINALIS DNA CONTROL
Catalogue Number MBC119-R
Lot Number 23MBC119001-R
Expiry date 2025-10-18

Kit contents	Lot Number	Expiry date
VIRCELL GIN DNA CONTROL	23MBC119101-R01	2025-10-18

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	20000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION


Identity test oligonucleotides

Giardia F GACGGCTCAGGACAACGGTT
Giardia R TTGCCAGCGGTGTCCG

Verweij, J.J., et al.(2003) Mol Cell Probes, 17(5):223-225

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Ramírez Galdeano, Jose Francisco	25/04/2023	



Vircell, S.L.
 Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264
 info@vircell.com www.vircell.com