

Certificate of Analysis

Product name	AMPLIRUN® ESCHERICHIA COLI (EAEC) DNA CONTROL
Catalogue Number	MBC121-R
Lot Number	24MBC121001-R
Expiry date	2025-11-05

Kit contents	Lot Number	Expiry date
VIRCELL EAEC DNA CONTROL	24MBC121101-R01	2025-11-05

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	15000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

EAEC-pCDV432-Fw CTGGCGAAAGACTGTATCAT
EAEC-pCDV432-Rv CAATGTATAGAAATCCGCTGTT

Bischoff, C., et al. (2005). J Microbiol Methods, 61(3): 335-341

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-R	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by

Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	06/06/2024	