

# Certificate of Analysis

**Product name** AMPLIRUN® ESCHERICHIA COLI (EIEC) DNA CONTROL  
**Catalogue Number** MBC122-R  
**Lot Number** 24MBC122001-R  
**Expiry date** 2027-05-27

Kit contents	Lot Number	Expiry date
VIRCELL EIEC DNA CONTROL	24MBC122101-R01	2027-05-27

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	11258 ± 790
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**  
*EIEC-ipaH Fw CT* TGGAAAACTCAGTGCCTCT  
*EIEC-ipaH Rv CT* CCAGTCCGTAAATTCATTCT  
*Bischoff, C., et al. (2005). J Microbiol Methods, 61(3): 335-341*

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.  
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	09/01/2025	