

# Certificate of Analysis

Product name	AMPLIRUN® ESCHERICHIA COLI (ETEC) DNA CONTROL
Catalogue Number	MBC124-R
Lot Number	24MBC124001-R
Expiry date	2025-09-28

Kit contents	Lot Number	Expiry date
VIRCELL ETEC DNA CONTROL	24MBC124101-R01	2025-09-28

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	18000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**

ETEC-elt.F CTCTATGTGCACACGGAGC

ETEC-elt.R CCATACTGATTGCCGCAAT

Hegde, A., et al. (2012). Indian J Med Microbiol, 30(3):279-284

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.  
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Martínez Ramírez, David	27/08/2024	