

# Certificate of Analysis

**Product name** AMPLIRUN® ESCHERICHIA COLI (ETEC) DNA CONTROL  
**Catalogue Number** MBC124-R  
**Lot Number** 23MBC124002-R  
**Expiry date** 2025-09-28


Kit contents	Lot Number	Expiry date
VIRCELL ETEC DNA CONTROL	23MBC124102-R01	2025-09-28

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	18000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**  
ETEC-elt.F CTCTATGTGCACACGGAGC  
ETEC-elt.R CCATACTGATTGCCGCAAT  
Hegde, A., et al. (2012). Indian J Med Microbiol, 30(3):279-284

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Ramírez Galdeano, Jose Francisco	30/03/2023	



Vircell, S.L.  
Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264  
info@vircell.com www.vircell.com