

Certificate of Analysis

Product name	AMPLIRUN® ENTEROVIRUS 68 RNA CONTROL
Catalogue Number	MBC125
Lot Number	25MBC125001-C
Expiry date	2027-02-19

Kit contents	Lot Number	Expiry date
VIRCELL ENT68 RNA CONTROL	25MBC125101-C01	2027-02-19

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	14777 ± 122
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides EV-D68Pavia-For GGCGGCCTACTCATGG EV-D68Pavia-Rev AGACTCTTCACACCTGTTCATGT <i>Piralla, A., et al. (2015). J Clin Microbiol, 53(5):1725-1726</i>		
--	--	--

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-C	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the applicable general safety and performance requirements as established in Regulation (EU) 2017/746. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	24/02/2025	