

Certificate of Analysis

Product name	AMPLIRUN® CRYPTOSPORIDIUM PARVUM DNA CONTROL
Catalogue Number	MBC126-R
Lot Number	25MBC126001-R
Expiry date	2026-04-03

Kit contents	Lot Number	Expiry date
VIRCELL CPA DNA CONTROL	25MBC126101-R01	2026-04-03

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	700 - 2000	1200
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:100 DILUTION	SPECIFIC AMPLIFICATION OF 1:100 DILUTION

Identity test oligonucleotides

CRyp.F CAAATTGATACCGTTTGTCTTCTG

CRyp.R GGCATGTCGATTCTAATTCAGCT

Guy, R. A., et al.(2003). Appl Environ Microbiol, 69(9):5178-5185

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	14/04/2025	