

# Certificate of Analysis

|                  |  |
|------------------|--|
| Product name     | AMPLIRUN® CRYPTOSPORIDIUM PARVUM DNA CONTROL |
| Catalogue Number | MBC126-R                                     |
| Lot Number       | 23MBC126001-R                                |
| Expiry date      | 2026-04-03                                   |

| Kit contents            | Lot Number      | Expiry date |
|-------------------------|-----------------|-------------|
| VIRCELL CPA DNA CONTROL | 23MBC126101-R01 | 2026-04-03  |

| Analysis                        | Specifications                           | Results                                  |
|---------------------------------|--|--|
| QUANTIFICATION TEST (copies/μL) | 700 - 2000                               | 1200                                     |
| IDENTITY TEST                   | SPECIFIC AMPLIFICATION OF 1:100 DILUTION | SPECIFIC AMPLIFICATION OF 1:100 DILUTION |


**Identity test oligonucleotides**

CRyp.F CAAATTGATACCGTTTGCCTTCTG  
 CRyp.R GGCATGTCGATTCTAATTCAGCT

Guy, R. A., et al.(2003). Appl Environ Microbiol, 69(9):5178-5185

|   |              |            |
|---|--------------|------------|
| VIRCELL CONTROL RECONSTITUTION SOLUTION | 22MBCRS002-R | 2027-11-16 |
|---|--------------|------------|

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

| Q.C. Testing by            |                         |            |   |
|----------------------------|-------------------------|------------|---|
| Position                   | Name                    | Date       | Signature   |
| Quality Control Department | Conde Ortega, Ana María | 06/10/2023 |  |



Vircell, S.L.  
 Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264  
 info@vircell.com www.vircell.com