

# Certificate of Analysis

**Product name** AMPLIRUN® ROSS RIVER VIRUS RNA CONTROL  
**Catalogue Number** MBC130-R  
**Lot Number** 24MBC130001-R  
**Expiry date** 2026-04-12

Kit contents	Lot Number	Expiry date
VIRCELL ROSS RIVER RNA CONTROL	24MBC130101-R01	2026-04-12

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	18883 ± 856
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**

RRV-E2int.F CGTTCGAGGACGCAGATTCA

RRV-E2int.R ATAGTCCTGCCGCTGCTGT

Soden, M., et al. (2000). Arthritis Rheum, 43(2):365-369

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-R	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.  
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	23/04/2024	