

Certificate of Analysis

Product name	AMPLIRUN® JAPANESE ENCEPHALITIS RNA CONTROL
Catalogue Number	MBC134-R
Lot Number	24MBC134001-R
Expiry date	2026-03-05

Kit contents	Lot Number	Expiry date
VIRCELL JEV RNA CONTROL	24MBC134101-R01	2026-03-05

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	18000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides Universal-JEV-F GCCACCCAGGAGGTCCTT Universal-JEV-R CCCCAAACCGCAGGAAT Shao, N., et al. (2018). Biomed and Environ Sci, 31(3):208-214		
--	--	--

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-R	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	18/03/2024	