

# Certificate of Analysis

Product name	AMPLIRUN® CORONAVIRUS OC43 RNA CONTROL
Catalogue Number	MBC135-R
Lot Number	24MBC135002-R
Expiry date	2026-09-04

Kit contents	Lot Number	Expiry date
VIRCELL OC43 RNA CONTROL	24MBC135102-R01	2026-09-04

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	14800 ± 404
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**

OC43-FP ATGTTAGGCCGATAATTGAGGACTAT  
 OC43-RP AATGTAAAGATGGCCGCGTATT

Vijgen, L., et al. (2005). J Clin Microbiol, 43(11):5452-5456

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.  
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	10/09/2024	