

Certificate of Analysis

Product name AMPLIRUN® SARS-CoV-2 RNA CONTROL
Catalogue Number MBC137
Lot Number 24MBC137002-C
Expiry date 2026-10-16

Kit contents	Lot Number	Expiry date
VIRCELL SARS-CoV-2 RNA CONTROL	24MBC137102-C01	2026-10-16

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	14711 ± 865
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

CorVWuh_E_F1 ACAGGTACGTTAATAGTTAATAGCGT

CorVWuh_E_R2 ATATTGCAGCAGTACGCACACA

Corman, V.M., et al. (2020) Euro Surveill., 25(3):2000045

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-C	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 and fulfils the essential requirements as established in Directive 98/79/CE and part of Regulation (EU) 2017/746.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	30/10/2024	