

# Certificate of Analysis

Product name	AMPLIRUN® SARS-CoV-2 P.1 RNA CONTROL
Catalogue Number	MBC140-R
Lot Number	24MBC140001-R
Expiry date	2026-10-08

Kit contents	Lot Number	Expiry date
VIRCELL SARS-CoV-2 P.1 RNA CONTROL	24MBC140101-R01	2026-10-08

  

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	18417 ± 704
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

  

**Identity test oligonucleotides**

CorVWuh\_E\_F1 ACAGGTACGTTAATAGTTAATAGCGT  
CorVWuh\_E\_R2 ATATTGCAGCAGTACGCACACA

Corman, V.M., et al. (2020) Euro Surveill., 25(3):2000045

  

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.  
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	11/10/2024	