

Certificate of Analysis

Product name	AMPLIRUN® SARS-CoV-2 OMICRON RNA CONTROL
Catalogue Number	MBC143-R
Lot Number	23MBC143001-R
Expiry date	2025-09-18

Kit contents	Lot Number	Expiry date
VIRCELL SARS-CoV-2 OMICRON RNA CONTROL	23MBC143101-R01	2025-09-18

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	13000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

CorVWuh_E_F1 ACAGGTACGTTAATAGTTAATAGCGT
CorVWuh_E_R2 ATATTGCAGCAGTACGCACACA

Corman, V.M., et al. (2020) Euro Surveill., 25(3):2000045

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Reca John, Isabel María	27/09/2023	