

Certificate of Analysis

Product name	AMPLIRUN® METAPNEUMOVIRUS RNA CONTROL
Catalogue Number	MBC144-R
Lot Number	24MBC144001-R
Expiry date	2026-04-29

Kit contents	Lot Number	Expiry date
VIRCELL hMPV RNA CONTROL	24MBC144101-R01	2026-04-29

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	19233 ± 724
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION
Identity test oligonucleotides HMPV-N-F2id GCACTCAAAAGATACCCTAGRATRGAYATAC HMPV-N-R2id AATAAACTTTCTGCTTGCTKCCTGTWGA Choudhary ML., et al. (2014) Arch Virol. 59(2):217-25		

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-R	2028-10-13
-----------------------------------------	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	16/05/2024	