

Certificate of Analysis

Product name AMPLIRUN® METAPNEUMOVIRUS RNA CONTROL
Catalogue Number MBC144-R
Lot Number 23MBC144001-R
Expiry date 2025-09-26


Kit contents	Lot Number	Expiry date
VIRCELL hMPV RNA CONTROL	23MBC144101-R01	2025-09-26

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	16000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides
HMPV-N-F2id GCACTCAAAAGATACCCTAGRATRGAYATAC
HMPV-N-R2id AATAAACTTTCTGCTTGCTKCCTGTWGA
Choudhary ML., et al. (2014) Arch Virol. 59(2):217-25

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	02/10/2023	



Vircell, S.L.
Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264
info@vircell.com www.vircell.com