

Certificate of Analysis

Product name	AMPLIRUN® SARS-CoV-2 OMICRON BA.2 RNA CONTROL
Catalogue Number	MBC145-R
Lot Number	24MBC145001-R
Expiry date	2025-12-18

Kit contents	Lot Number	Expiry date
VIRCELL SARS-CoV-2 OMICRON BA.2 RNA CONTROL	24MBC145101-R01	2025-12-18

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	14000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

CorVWuh_E_F1 ACAGGTACGTTAATAGTTAATAGCGT
CorVWuh_E_R2 ATATTGCAGCAGTACGCACACA

Corman, V.M., et al. (2020) Euro Surveill., 25(3):2000045

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-R	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	27/05/2024	