

Certificate of Analysis

| | |
|------------------|---|
| Product name | AMPLIRUN® SARS-CoV-2 OMICRON BA.2 RNA CONTROL |
| Catalogue Number | MBC145-R |
| Lot Number | 23MBC145001-R |
| Expiry date | 2025-12-18 |

| Kit contents | Lot Number | Expiry date |
|---|-----------------|-------------|
| VIRCELL SARS-CoV-2 OMICRON BA.2 RNA CONTROL | 23MBC145101-R01 | 2025-12-18 |

| Analysis | Specifications | Results |
|---------------------------------|---|---|
| QUANTIFICATION TEST (copies/μL) | 12500 - 20000 | 14000 |
| IDENTITY TEST | SPECIFIC AMPLIFICATION OF 1:1000 DILUTION | SPECIFIC AMPLIFICATION OF 1:1000 DILUTION |

Identity test oligonucleotides

CorVWuh_E_F1 ACAGGTACGTTAATAGTTAATAGCGT
CorVWuh_E_R2 ATATTGCAGCAGTACGCACACA

Corman, V.M., et al. (2020) Euro Surveill., 25(3):2000045

| | | |
|---|--------------|------------|
| VIRCELL CONTROL RECONSTITUTION SOLUTION | 23MBCRS001-R | 2028-10-13 |
|---|--------------|------------|

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by

| Position | Name | Date | Signature |
|----------------------------|-------------------------|------------|---|
| Quality Control Department | Conde Ortega, Ana María | 26/12/2023 |  |