

Certificate of Analysis

Product name AMPLIRUN® ASTROVIRUS RNA CONTROL
Catalogue Number MBC147-R
Lot Number 23MBC147001-R
Expiry date 2025-07-19

Kit contents	Lot Number	Expiry date
VIRCELL AST RNA CONTROL	23MBC147101-R01	2025-07-19

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	18000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION


Identity test oligonucleotides

AstV_ORF1_F AATTGCTCAAAGAGGAAATAGA
 AstV_ORF1_R GCAAGTATVCCATTGATTCAT

Li, W., Li, W., Li, L. et al. Multiplex detection of eight different viral enteropathogens in clinical samples, combining RT-PCR technology with melting curve analysis. *Viral J* 19, 61 (2022). <https://doi.org/10.1186/s12985-022-01789-z>

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	22/09/2023	



Vircell, S.L.
 Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264
 info@vircell.com www.vircell.com