

Certificate of Analysis

Product name	AMPLIRUN® OROPOUCHE VIRUS RNA CONTROL
Catalogue Number	MBC152-R
Lot Number	24MBC152001-R
Expiry date	2026-09-17

Kit contents	Lot Number	Expiry date
VIRCELL ORV RNA CONTROL	24MBC152101-R01	2026-09-17

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	18720 ± 459
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

OROV_FNF
TCCGGAGGCAGCATATGTG
OROV_FNR
ACAACACCAGCATTGAGCACTT

https://www.researchgate.net/publication/318081666_Multiplexed_reverse_transcription_real-time_polymerase_chain_reaction_for_simultaneous_detection_of_Mayaro_Oropouche_and_Oropouche-like_viruses

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	05/11/2024	