

Certificate of Analysis

Product name	AMPLIRUN® TOTAL MTB CONTROL (SPUTUM)
Catalogue Number	MBTC013
Lot Number	24MBTC013001-C
Expiry date	2026-12-18

Kit contents	Lot Number	Expiry date
VIRCELL TOTAL MTB CONTROL (SPUTUM)	24MBTC013103-C	2026-12-18

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/vial)	20000-50000	37000
IDENTITY TEST	SPECIFIC AMPLIFICATION	SPECIFIC AMPLIFICATION
Identity test oligonucleotides IS6110f TCAAGGAGCACATCAGCC IS6110r TCACGGTTCAGGGTTAGC Sanjuan-Jimenez R., et al. (2013). PLoS One, 8(3): e58353		

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	25/06/2024	