

# Certificate of Analysis

Product name	AMPLIRUN® TOTAL MTB CONTROL (SPUTUM)
Catalogue Number	MBTC013
Lot Number	25MBTC013001-C
Expiry date	2027-09-25

Kit contents	Lot Number	Expiry date
VIRCELL TOTAL MTB CONTROL (SPUTUM)	25MBTC013101-C	2027-09-25

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/vial)	20000-50000	45000
IDENTITY TEST	SPECIFIC AMPLIFICATION	SPECIFIC AMPLIFICATION
<b>Identity test oligonucleotides</b> IS6110f TCAAGGAGCACATCAGCC IS6110r TCACGGTTCAGGGTTAGC Sanjuan-Jimenez R., et al. (2013). PLoS One, 8(3): e58353		

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE and part of Regulation (EU) 2017/746. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

## Q.C. Testing by

Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	27/03/2025	