

Certificate of Analysis

Product name	AMPLIRUN® TOTAL MTB RIF RESISTANT CONTROL (SPUTUM)
Catalogue Number	MBTC014
Lot Number	24MBTC014002-C
Expiry date	2026-12-11

Kit contents	Lot Number	Expiry date
VIRCELL TOTAL MTB RIF RESISTANT (531) CONTROL (SPUTUM)	24MBTC014104-C	2026-12-11

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/vial)	20000-50000	34000
IDENTITY TEST	SPECIFIC AMPLIFICATION	SPECIFIC AMPLIFICATION
Identity test oligonucleotides <i>IS6110f</i> TCAAGGAGCACATCAGCC <i>IS6110r</i> TCACGGTTCAGGGTTAGC <i>Sanjuan-Jimenez R., et al. (2013). PLoS One, 8(3): e58353</i>		

VIRCELL TOTAL MTB RIF RESISTANT (526) CONTROL (SPUTUM)	24MBTC014204-C	2027-01-24
--	----------------	------------

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/vial)	20000-50000	25000
IDENTITY TEST	SPECIFIC AMPLIFICATION	SPECIFIC AMPLIFICATION
Identity test oligonucleotides <i>IS6110f</i> TCAAGGAGCACATCAGCC <i>IS6110r</i> TCACGGTTCAGGGTTAGC <i>Sanjuan-Jimenez R., et al. (2013). PLoS One, 8(3): e58353</i>		

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE and part of Regulation (EU) 2017/746. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by

Position	Name	Date	Signature
Quality Control Department	Martínez Ramírez, David	30/07/2024	